



**CERTIFICATE OF ANALYSIS**

Nova 8 Calibrator Pack  
Lot Number 808763

Ref Number 15196  
Exp Date 2019-09-20  
Date of Manufacture 2018-03-20  
UDI# (01)00385480151962(11)180320(17)190920(10)808763

Cal A L/N	18079031	Spec	Pass/Fail	Cal B L/N	18085025	Spec	Pass/Fail	Cal C L/N	18079033	Spec	Pass/Fail
Na mmol/L	139.0-141.0		Pass	Na mmol/L	74.5-75.5		Pass	Na mmol/L	139.0-141.0		Pass
K mmol/L	3.97-4.03		Pass	K mmol/L	9.8-10.2		Pass	K mmol/L	3.97-4.03		Pass
iCa mmol/L	0.99-1.01		Pass	iCa mmol/L	1.98-2.02		Pass	iCa mmol/L	0.99-1.01		Pass
iMg mmol/L	0.495-0.505		Pass	iMg mmol/L	0.495-0.505		Pass	iMg mmol/L	1.48-1.52		Pass
pH @ 37C	7.274-7.294		Pass	pH @ 37C	6.730-6.750		Pass	pH @ 37C	7.374-7.394		Pass

  

Reference L/N	18081037	Spec	Pass/Fail
KCl mol/L	1.8-2.2		Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and ISO EN 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Inspector  
Title

4/2/2018  
Date