



**CERTIFICATE OF ANALYSIS**

Nova 8 Calibrator Pack  
Lot Number 823229

Ref Number 15196  
Exp Date 2020-02-06  
Date of Manufacture 2018-08-06  
UDI# (01)00385480151962(11)180806(17)200206(10)823229

Cal A	L/N	18222025	Spec	Pass/Fail	Cal B	L/N	18222026	Spec	Pass/Fail	Cal C	L/N	18222027	Spec	Pass/Fail
Na mmol/L	139.0-141.0	Pass	Na mmol/L	74.5-75.5	Pass	Na mmol/L	139.0-141.0	Pass						
K mmol/L	3.97-4.03	Pass	K mmol/L	9.8-10.2	Pass	K mmol/L	3.97-4.03	Pass						
iCa mmol/L	0.99-1.01	Pass	iCa mmol/L	1.98-2.02	Pass	iCa mmol/L	0.99-1.01	Pass						
iMg mmol/L	0.495-0.505	Pass	iMg mmol/L	0.495-0.505	Pass	iMg mmol/L	1.48-1.52	Pass						
pH @ 37C	7.274-7.294	Pass	pH @ 37C	6.730-6.750	Pass	pH @ 37C	7.374-7.394	Pass						
					Reference	L/N	18218007	Spec	Pass/Fail					
					KCl mol/L	1.8-2.2	Pass							

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and ISO EN 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Inspector  
Title

8/28/2018  
Date