

ABX Minotrol CRP



LOT

MC 417

CONTROL



(Exp.) 2021-01-05
(YYYY-MM-DD)

Rev 1

| PARAMETRES PARAMETERS | | UNITES UNITS | | ABX Lysebio | | | | | | | | | | | | | | | |
|--------------------------|------|---|--------------|-------------|--------------|--------|---------|-------------------------|---------|--------------|--------|---------|-------------------------|---------|--------------|--------|---------|-------------------------|-------------------------|
| | | | | CONTROL | | | | 1 | CONTROL | | | | 2 | CONTROL | | | | 3 | TOLERANCES TOLERANCE |
| | | | | MICROS | | PENTRA | | TOLERANCES TOLERANCE | MICROS | | PENTRA | | TOLERANCES TOLERANCE | MICROS | | PENTRA | | TOLERANCES TOLERANCE | |
| | | | | EMI CRP | | MS CRP | | | EMI CRP | | MS CRP | | | EMI CRP | | MS CRP | | | |
| GB | WBC | 10 ³ /mm ³ ;10 ⁹ /l | 2.0 | | 2.0 | | ± 0.4 | 7.7 | | 7.9 | | ± 1.0 | 20.1 | | 20.9 | | ± 2.2 | | |
| GR | RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 2.43 | | 2.40 | | ± 0.15 | 4.55 | | 4.55 | | ± 0.18 | 5.42 | | 5.45 | | ± 0.25 | | |
| HB | HGB | g/dl | 6.2 | | 5.9 | | ± 0.4 | 13.4 | | 13.5 | | ± 0.5 | 17.6 | | 17.9 | | ± 0.7 | | |
| | | g/l | 62 | | 59 | | ± 4 | 134 | | 135 | | ± 5 | 176 | | 179 | | ± 7 | | |
| | | mmol/l | 3.85 | | 3.66 | | ± 0.25 | 8.32 | | 8.38 | | ± 0.31 | 10.93 | | 11.12 | | ± 0.44 | | |
| HT | HCT | % | 17.0 | | 16.6 | | ± 2.0 | 37.3 | | 36.9 | | ± 2.5 | 48.8 | | 48.5 | | ± 3.0 | | |
| | | l/l | 0.170 | | 0.166 | | ± 0.020 | 0.373 | | 0.369 | | ± 0.025 | 0.488 | | 0.485 | | ± 0.030 | | |
| VGM | MCV | µm ³ ; fl | 70 | | 69.0 | | ± 4 | 82 | | 81.0 | | ± 4 | 90 | | 89.0 | | ± 4 | | |
| TGMH | MCH | pg | 25.5 | | 24.6 | | ± 2.4 | 29.5 | | 29.7 | | ± 2.4 | 32.5 | | 32.8 | | ± 2.8 | | |
| | | fmol | 1.58 | | 1.53 | | ± 0.15 | 1.83 | | 1.84 | | ± 0.15 | 2.02 | | 2.04 | | ± 0.18 | | |
| CCMH | MCHC | g/dl | 36.4 | | 35.6 | | ± 3.0 | 35.9 | | 36.6 | | ± 3.0 | 36.1 | | 36.9 | | ± 3.0 | | |
| | | g/l | 364 | | 356 | | ± 30 | 359 | | 366 | | ± 30 | 361 | | 369 | | ± 30 | | |
| | | mmol/l | 22.63 | | 22.13 | | ± 1.86 | 22.30 | | 22.75 | | ± 1.86 | 22.41 | | 22.92 | | ± 1.86 | | |
| IDR | RDW | % | 14.0 | | 14.5 | | ± 3.0 | 13.0 | | 13.2 | | ± 3.0 | 12.0 | | 11.8 | | ± 3.0 | | |
| PLAQ. | PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 68 | | 67 | | ± 20 | 251 | | 255 | | ± 40 | 489 | | 523 | | ± 60 | | |
| VPM | MPV | µm ³ ; fl | 8.2 | | 9.5 | | ± 2.0 | 8.4 | | 9.2 | | ± 2.0 | 8.1 | | 8.4 | | ± 2.0 | | |
| LYMPHO | | # | 1.3 | | N/A | | ± 0.2 | 2.8 | | N/A | | ± 0.5 | 3.5 | | N/A | | ± 1.2 | | |
| | | % | 64.5 | | N/A | | ± 8.0 | 36.0 | | N/A | | ± 6.0 | 17.5 | | N/A | | ± 6.0 | | |
| MONO | | # | 0.2 | | N/A | | ± 0.2 | 0.5 | | N/A | | ± 0.4 | 0.9 | | N/A | | ± 0.9 | | |
| | | % | 8.0 | | N/A | | ± 7.0 | 6.5 | | N/A | | ± 5.0 | 4.5 | | N/A | | ± 4.5 | | |
| GRANULO | | # | 0.6 | | N/A | | ± 0.2 | 4.4 | | N/A | | ± 0.5 | 15.7 | | N/A | | ± 1.5 | | |
| | | % | 27.5 | | N/A | | ± 7.0 | 57.5 | | N/A | | ± 6.0 | 78.0 | | N/A | | ± 7.0 | | |
| CRP | | mg/dl | 0.60 | | 0.56 | | ± 0.40 | 1.95 | | 1.86 | | ± 0.80 | 4.10 | | 3.97 | | ± 1.50 | | |
| | | mg/l | 6.0 | | 5.6 | | ± 4.0 | 19.5 | | 18.6 | | ± 8.0 | 41.0 | | 39.7 | | ± 15.0 | | |

Ref: TEMP-1070 Rev.38 BACK/VERSO 1300060514