

ABX Minotrol CRP



LOT

MC 426

CONTROL



(Exp.) 2023-04-05

(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio													
				CONTROL		1	CONTROL		2	CONTROL		3	CONTROL		TOLERANCES TOLERANCE		
				MICROS		PENTRA		MICROS		PENTRA		MICROS		PENTRA			
				EMI CRP	LC-767G	MS CRP		EMI CRP	LC-767G	MS CRP		EMI CRP	LC-767G	MS CRP			
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0	2.1		± 0.4	7.7	7.7	8.0		± 1.0	20.3	20.8		21.6	
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.35	2.37	2.35		± 0.15	4.45	4.46	4.50		± 0.18	5.37	5.38	5.49		± 0.25
HB	HGB	g/dl	6.3	6.3	6.0		± 0.4	13.3	13.4	13.4		± 0.5	17.3	17.4	17.8		± 0.7
		g/l	63	63	60		± 4	133	134	134		± 5	173	174	178		± 7
		mmol/l	3.91	3.91	3.73		± 0.25	8.26	8.32	8.32		± 0.31	10.74	10.81	11.05		± 0.44
HT	HCT	%	16.9	17.1	16.8		± 2.0	36.5	36.6	36.5		± 2.5	47.8	47.9	48.3		± 3.0
		l/l	0.169	0.171	0.168		± 0.020	0.365	0.366	0.365		± 0.025	0.478	0.479	0.483		± 0.030
VGM	MCV	µm ³ ; fl	72	72	71.5		± 4	82	82	81.0		± 4	89	89	88.0		± 4
TGMH	MCH	pg	26.8	26.6	25.5		± 2.4	29.9	30.0	29.8		± 2.4	32.2	32.3	32.4		± 2.8
		fmol	1.66	1.65	1.59		± 0.15	1.86	1.87	1.85		± 0.15	2.00	2.01	2.01		± 0.18
CCMH	MCHC	g/dl	37.2	36.9	35.7		± 3.0	36.4	36.6	36.8		± 3.0	36.2	36.3	36.8		± 3.0
		g/l	372	369	357		± 30	364	366	368		± 30	362	363	368		± 30
		mmol/l	23.12	22.93	22.18		± 1.86	22.63	22.75	22.83		± 1.86	22.48	22.57	22.88		± 1.86
IDR	RDW	%	14.0	13.5	15.0		± 3.0	13.0	13.0	13.0		± 3.0	12.5	12.5	12.5		± 3.0
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	72	71	64		± 20	270	270	267		± 40	513	511	522		± 60
VPM	MPV	µm ³ ; fl	8.0	8.8	9.7		± 2.0	8.4	8.6	9.1		± 2.0	8.6	8.6	9.1		± 2.0
LYMPHO		#	1.3	1.3	N/A		± 0.2	2.6	2.5	N/A		± 0.5	3.5	3.3	N/A		± 1.1
		%	63.5	63.0	N/A		± 8.0	33.5	33.0	N/A		± 6.0	17.0	16.0	N/A		± 5.0
MONO		#	0.1	0.2	N/A		± 0.2	0.5	0.6	N/A		± 0.4	1.2	1.2	N/A		± 1.0
		%	7.0	7.5	N/A		± 7.0	7.0	7.5	N/A		± 5.0	6.0	6.0	N/A		± 5.0
GRANULO		#	0.6	0.6	N/A		± 0.2	4.6	4.6	N/A		± 0.5	15.6	16.2	N/A		± 1.5
		%	29.5	29.5	N/A		± 7.0	59.5	59.5	N/A		± 6.0	77.0	78.0	N/A		± 7.0
CRP		mg/dl	0.47	0.53	0.52		± 0.40	2.17	2.32	2.13		± 0.80	4.52	4.72	4.61		± 1.50
		mg/l	4.7	5.3	5.2		± 4.0	21.7	23.2	21.3		± 8.0	45.2	47.2	46.1		± 15.0

Ref: TEMP-1070 Rev.42 BACKVERSO 1300092170