

# ABX Minotrol CRP



**LOT**

**MC 431**

**CONTROL**



**(Exp.) 2024-07-05**  
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio													
				CONTROL			1	CONTROL			2	CONTROL			3	TOLERANCES TOLERANCE	
				MICROS		PENTRA	TOLERANCES TOLERANCE	MICROS		PENTRA	TOLERANCES TOLERANCE	MICROS		PENTRA			
				EMI CRP	LC-767G	MS CRP		EMI CRP	LC-767G	MS CRP		EMI CRP	LC-767G	MS CRP			
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	2.2	2.1	2.1	± 0.4	8.0		7.9	8.1	± 1.0		21.3	21.4	22.1	± 2.2		
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	2.30	2.30	2.20	± 0.15	4.50	4.48	4.60	± 0.18	5.50	5.45	5.60	± 0.25				
HB HGB	g/dl	6.3	6.2	6.0	± 0.4	13.2	13.2	13.4	± 0.5	17.6	17.5	18.1	± 0.7				
	g/l	63	62	60	± 4	132	132	134	± 5	176	175	181	± 7				
	mmol/l	3.91	3.85	3.73	± 0.25	8.20	8.20	8.32	± 0.31	10.93	10.87	11.24	± 0.44				
HT HCT	%	16.8	16.8	16.4	± 2.0	36.0	35.8	36.8	± 2.5	48.4	48.0	49.0	± 3.0				
	l/l	0.168	0.168	0.164	± 0.020	0.360	0.358	0.368	± 0.025	0.484	0.480	0.490	± 0.030				
VGM MCV	µm <sup>3</sup> ; fl	73	73	74.5	± 4	80	80	80.0	± 4	88	88	87.5	± 4				
TGMH MCH	pg	27.4	27.0	27.3	± 2.4	29.3	29.5	29.1	± 2.4	32.0	32.1	32.3	± 2.8				
	fmol	1.70	1.67	1.69	± 0.15	1.82	1.83	1.81	± 0.15	1.99	1.99	2.01	± 0.18				
CCMH MCHC	g/dl	37.5	36.9	36.6	± 3.0	36.7	36.8	36.4	± 3.0	36.4	36.5	36.9	± 3.0				
	g/l	375	369	366	± 30	367	368	364	± 30	364	365	369	± 30				
	mmol/l	23.30	22.93	22.73	± 1.86	22.77	22.87	22.61	± 1.86	22.58	22.66	22.94	± 1.86				
IDR RDW	%	13.5	13.0	15.5	± 3.0	14.0	13.5	13.0	± 3.0	13.5	13.0	12.0	± 3.0				
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	77	73	73	± 20	256	257	271	± 40	500	509	527	± 60				
VPM MPV	µm <sup>3</sup> ; fl	7.5	8.4	8.9	± 2.0	7.5	7.8	8.3	± 2.0	7.7	7.9	8.4	± 2.0				
LYMPHO	#	1.4	1.3	N/A	± 0.2	2.8	2.7	N/A	± 0.5	3.6	3.4	N/A	± 1.1				
	%	64.5	64.0	N/A	± 8.0	35.0	34.5	N/A	± 6.0	17.0	16.0	N/A	± 5.0				
MONO	#	0.2	0.2	N/A	± 0.2	1.0	0.8	N/A	± 0.4	2.2	1.7	N/A	± 1.0				
	%	8.0	8.0	N/A	± 7.0	12.0	9.5	N/A	± 5.0	10.5	8.0	N/A	± 5.0				
GRANULO	#	0.6	0.6	N/A	± 0.2	4.2	4.4	N/A	± 0.5	15.4	16.3	N/A	± 1.5				
	%	27.5	28.0	N/A	± 7.0	53.0	56.0	N/A	± 6.0	72.5	76.0	N/A	± 7.0				
CRP	mg/dl	0.54	0.51	0.55	± 0.40	2.52	2.60	2.63	± 0.80	4.75	4.88	5.07	± 1.50				
	mg/l	5.4	5.1	5.5	± 4.0	25.2	26.0	26.3	± 8.0	47.5	48.8	50.7	± 15.0				

Ref: TEMP-1070 Rev.42 BACKVERSO 1300092170