

ABX Minotrol CRP


LOT
MC 437
CONTROL

(Exp.) 2026-01-05

(YYYY - MM - DD)

Rev 1

ABX Lysebio

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	
			CONTROL				1	TOLERANCES TOLERANCE	CONTROL				2	TOLERANCES TOLERANCE	CONTROL				3	TOLERANCES TOLERANCE
			MICROS			PENTRA			MICROS			PENTRA			MICROS			PENTRA		
			EMI CRP	LC-767G		MS CRP			EMI CRP	LC-767G		MS CRP			EMI CRP	LC-767G		MS CRP		
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	1.9		2.0		± 0.4	7.9	7.8		8.1		± 1.0	20.0	20.1		20.9		± 2.2
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.34	2.32		2.29		± 0.15	4.53	4.46		4.51		± 0.18	5.56	5.47		5.60		± 0.25
HB	HGB	g/dl	6.1	6.0		5.8		± 0.4	13.2	13.3		13.4		± 0.5	17.4	17.5		18.0		± 0.7
		g/l	61	60		58		± 4	132	133		134		± 5	174	175		180		± 7
		mmol/l	3.79	3.73		3.60		± 0.25	8.20	8.26		8.32		± 0.31	10.81	10.87		11.18		± 0.44
HT	HCT	%	16.4	16.2		16.3		± 2.0	36.7	36.6		36.8		± 2.5	48.9	48.1		49.0		± 3.0
		l/l	0.164	0.162		0.163		± 0.020	0.367	0.366		0.368		± 0.025	0.489	0.481		0.490		± 0.030
VGM	MCV	µm ³ ; fl	70	70		71.0		± 4	81	82		81.5		± 4	88	88		87.5		± 4
TGMH	MCH	pg	26.1	25.9		25.3		± 2.4	29.1	29.8		29.7		± 2.4	31.3	32.0		32.1		± 2.8
		fmol	1.62	1.61		1.57		± 0.15	1.81	1.85		1.85		± 0.15	1.94	1.99		2.00		± 0.18
CCMH	MCHC	g/dl	37.2	36.9		35.7		± 3.0	36.0	36.4		36.5		± 3.0	35.6	36.4		36.7		± 3.0
		g/l	372	369		357		± 30	360	364		365		± 30	356	364		367		± 30
		mmol/l	23.13	22.94		22.15		± 1.86	22.34	22.58		22.64		± 1.86	22.08	22.58		22.81		± 1.86
IDR	RDW	%	15.0	14.5		16.0		± 3.0	14.5	13.5		13.0		± 3.0	14.5	13.5		12.5		± 3.0
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	72	76		62		± 20	249	249		235		± 40	489	491		481		± 60
VPM	MPV	µm ³ ; fl	8.4	9.3		9.8		± 2.0	8.4	8.7		9.3		± 2.0	8.6	8.6		9.1		± 2.0
LYMPHO	#	1.3	1.2		N/A		± 0.2	2.5	2.5		N/A		± 0.5	3.1	3.0		N/A		± 1.1	
	%	64.0	64.0		N/A		± 8.0	31.5	31.5		N/A		± 6.0	15.5	15.0		N/A		± 5.0	
MONO	#	0.2	0.2		N/A		± 0.2	0.5	0.5		N/A		± 0.5	1.2	1.1		N/A		± 1.0	
	%	8.0	8.5		N/A		± 7.2	6.5	7.0		N/A		± 5.9	6.0	5.5		N/A		± 5.0	
GRANULO	#	0.6	0.5		N/A		± 0.2	4.9	4.8		N/A		± 0.8	15.7	16.0		N/A		± 2.0	
	%	28.0	27.5		N/A		± 9.9	62.0	61.5		N/A		± 9.9	78.5	79.5		N/A		± 9.9	
CRP	mg/dl	0.51	0.58		0.54		± 0.40	2.79	2.99		2.81		± 0.80	5.13	5.47		5.41		± 1.50	
	mg/l	5.1	5.8		5.4		± 4.0	27.9	29.9		28.1		± 8.0	51.3	54.7		54.1		± 15.0	

Ref: TEMP-1070 Rev.43 BACK/VERSO 1300092170