

ABX Minotrol CRP



LOT

MC 439

CONTROL



(Exp.) 2026-07-05
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio																
				CONTROL				1	CONTROL				2	CONTROL				3		
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE		
				EMI CRP	LC-767G		MS CRP			EMI CRP	LC-767G			MS CRP		EMI CRP	LC-767G			MS CRP
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0		2.0		± 0.4	7.7	7.6		7.7		± 1.0	20.1	20.3		20.7		± 2.2
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.50	2.50		2.41		± 0.15	4.83	4.74		4.75		± 0.18	5.74	5.73		5.75		± 0.25
HB	HGB	g/dl	6.4	6.4		6.2		± 0.4	13.9	13.9		14.1		± 0.5	18.2	18.3		18.7		± 0.7
		g/l	64	64		62		± 4	139	139		141		± 5	182	183		187		± 7
		mmol/l	3.97	3.97		3.85		± 0.25	8.63	8.63		8.76		± 0.31	11.30	11.36		11.61		± 0.44
HT	HCT	%	17.5	17.3		17.1		± 2.0	38.6	37.9		38.0		± 2.5	51.1	50.4		50.0		± 3.0
		l/l	0.175	0.173		0.171		± 0.020	0.386	0.379		0.380		± 0.025	0.511	0.504		0.500		± 0.030
VGM	MCV	µm ³ ; fl	70	69		71.0		± 4	80	80		80.0		± 4	89	88		87.0		± 4
TGMH	MCH	pg	25.6	25.6		25.7		± 2.4	28.8	29.3		29.7		± 2.4	31.7	31.9		32.5		± 2.8
		fmol	1.59	1.59		1.60		± 0.15	1.79	1.82		1.84		± 0.15	1.97	1.98		2.02		± 0.18
CCMH	MCHC	g/dl	36.6	37.1		36.2		± 3.0	36.0	36.7		37.1		± 3.0	35.6	36.3		37.4		± 3.0
		g/l	366	371		362		± 30	360	367		371		± 30	356	363		374		± 30
		mmol/l	22.71	23.04		22.50		± 1.86	22.34	22.76		23.04		± 1.86	22.12	22.54		23.21		± 1.86
IDR	RDW	%	13.5	14.0		14.5		± 3.0	14.0	13.5		12.5		± 3.0	14.0	13.5		11.5		± 3.0
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	71	74		65		± 20	244	243		246		± 40	491	485		490		± 60
VPM	MPV	µm ³ ; fl	7.5	8.4		9.5		± 2.0	7.3	7.6		8.7		± 2.0	7.8	8.1		8.8		± 2.0
LYMPHO		#	1.2	1.2		N/A		± 0.2	2.5	2.4		N/A		± 0.5	3.1	3.0		N/A		± 1.1
		%	62.0	61.0		N/A		± 8.0	32.5	32.0		N/A		± 6.0	15.5	15.0		N/A		± 5.0
MONO		#	0.2	0.2		N/A		± 0.2	0.6	0.6		N/A		± 0.5	1.9	1.8		N/A		± 1.6
		%	9.0	9.5		N/A		± 8.1	7.5	7.5		N/A		± 6.8	9.5	9.0		N/A		± 8.1
GRANULO		#	0.6	0.6		N/A		± 0.2	4.6	4.6		N/A		± 0.8	15.1	15.4		N/A		± 2.0
		%	29.0	29.5		N/A		± 9.9	60.0	60.5		N/A		± 9.9	75.0	76.0		N/A		± 9.9
CRP		mg/dl	0.88	0.89		0.91		± 0.40	2.20	2.19		2.19		± 0.80	3.97	3.89		4.19		± 1.50
		mg/l	8.8	8.9		9.1		± 4.0	22.0	21.9		21.9		± 8.0	39.7	38.9		41.9		± 15.0

Ref: TEMP-1070 Rev.43 BACKVERSO 1300092170