

ABX Minotrol CRP


LOT
MC 441

Rev 1

CONTROL

 (Exp.) **2027-01-05**

(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																			
			CONTROL					1	CONTROL					2		CONTROL					3	TOLERANCES TOLERANCE
			MICROS			PENTRA		TOLERANCES TOLERANCE	MICROS			PENTRA		TOLERANCES TOLERANCE	MICROS			PENTRA				
			EMI CRP	LC-767G		MS CRP			EMI CRP	LC-767G		MS CRP			EMI CRP	LC-767G		MS CRP				
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.2	2.2		2.1		± 0.4	7.8	7.9		7.9		± 1.0	20.3	20.4		21.0		± 2.2		
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.22	2.24		2.18		± 0.15	4.42	4.41		4.45		± 0.18	5.64	5.64		5.71		± 0.25		
HB	HGB	g/dl	6.1	6.1		5.8		± 0.4	13.3	13.3		13.5		± 0.5	17.7	17.9		18.3		± 0.7		
		g/l	61	61		58		± 4	133	133		135		± 5	177	179		183		± 7		
		mmol/l	3.79	3.79		3.60		± 0.25	8.26	8.26		8.38		± 0.31	10.99	11.12		11.36		± 0.44		
HT	HCT	%	16.4	16.4		15.9		± 2.0	36.7	36.2		36.5		± 2.5	49.6	49.1		48.8		± 3.0		
		l/l	0.164	0.164		0.159		± 0.020	0.367	0.362		0.365		± 0.025	0.496	0.491		0.488		± 0.030		
VGM	MCV	µm ³ ; fl	74	73		73.0		± 4	83	82		82.0		± 4	88	87		85.5		± 4		
TGMH	MCH	pg	27.5	27.2		26.6		± 2.4	30.1	30.2		30.3		± 2.4	31.4	31.7		32.0		± 2.8		
		fmol	1.71	1.69		1.65		± 0.15	1.87	1.87		1.88		± 0.15	1.95	1.97		1.99		± 0.18		
CCMH	MCHC	g/dl	37.1	37.3		36.4		± 3.0	36.3	36.8		37.0		± 3.0	35.7	36.5		37.5		± 3.0		
		g/l	371	373		364		± 30	363	368		370		± 30	357	365		375		± 30		
		mmol/l	23.06	23.17		22.63		± 1.86	22.51	22.84		22.97		± 1.86	22.15	22.65		23.28		± 1.86		
IDR	RDW	%	14.0	14.0		14.5		± 3.0	13.5	13.0		13.0		± 3.0	14.0	13.0		11.5		± 3.0		
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	65	72		65		± 20	242	245		243		± 40	469	486		497		± 60		
VPM	MPV	µm ³ ; fl	7.4	8.1		8.7		± 2.0	7.8	8.0		8.5		± 2.0	7.6	7.9		7.9		± 2.0		
LYMPHO	#		1.4	1.3		N/A		± 0.2	2.4	2.4		N/A		± 0.5	3.1	3.0		N/A		± 1.1		
	%		61.5	61.0		N/A		± 8.0	31.0	31.0		N/A		± 6.0	15.5	14.5		N/A		± 5.0		
MONO	#		0.2	0.2		N/A		± 0.2	0.8	0.8		N/A		± 0.7	1.4	1.3		N/A		± 1.2		
	%		8.5	8.5		N/A		± 7.7	10.5	10.0		N/A		± 9.0	7.0	6.5		N/A		± 5.9		
GRANULO	#		0.7	0.7		N/A		± 0.2	4.6	4.7		N/A		± 0.8	15.7	16.1		N/A		± 2.0		
	%		30.0	30.5		N/A		± 9.9	58.5	59.0		N/A		± 9.9	77.5	79.0		N/A		± 9.9		
CRP	mg/dl		0.93	0.94		0.88		± 0.40	2.50	2.64		2.29		± 0.80	4.27	4.46		4.12		± 1.50		
	mg/l		9.3	9.4		8.8		± 4.0	25.0	26.4		22.9		± 8.0	42.7	44.6		41.2		± 15.0		

 1300092170
 BACK/VERSO
 Ref: TEMP-1070 Rev.43